IOWA DEPARTMENT OF NATURAL RESOURCES

GOVERNOR KIM REYNOLDS
LT. GOVERNOR ADAM GREGG
ACTING DIRECTOR BRUCE TRAUTMAN

April 11, 2019

City of North Liberty
PO Box 77
13909 South Front Street
North Liberty IA 52317

SUBJECT: Incorrect Total Kjeldahl Nitrogen (TKN) results
City of North Liberty – NPDES Permit #5252001

City Officials:

The State Hygienic Laboratory (Ankeny Lab #27) has notified the Iowa DNR of incorrect Total Kjeldahl Nitrogen (TKN) results that were reported from October 31, 2018 through January 18, 2019. The correct results have been recalculated. In all cases, the corrected results are lower than the original results by a factor of 0.824. A spreadsheet of the results you received and the corrected results is attached.

Action Required:

- You must review the data included in these DMRs and determine if there are inaccuracies.
- If it is discovered that the data on the previously submitted DMRs is inaccurate, a revised DMR must be resubmitted to the Department no later than 30-days from the date of this letter.

The Iowa DNR acknowledges that your facility complied with the monitoring and reporting requirements of the NPDES permit during the above mentioned time period. Please continue to sample as per your NPDES Permit.

If you have questions or comments, please do not hesitate to contact Terry Jones in this office at 319-653-2135. Thank you for your understanding and cooperation regarding this matter.

Sincerely,
FIELD SERVICES & COMPLIANCE BUREAU

Terry Jones

for Deborah Quade
Supervisor, Field Office #6

ESD06:N\tjones\ww\north liberty\tkn ltr NL 4-2019

Encl. Lab Analysis Results
xc: DNR Records Section, DNR, Des Moines
North Liberty WW File